



Instrument Kit

Membrane Standard kit

INSTRUCTIONS FOR USE

DESCRIPTION:

The device is intended for use in inserting and removing dental bone screws for guided bone regeneration in dental implant surgery.

COMPOSITION:



GBR Screws Motor Mount/ Driver: MMGB20 (Ø2.3 X 20) / MMGB26 (Ø2.3 X 26).

Mallet: GMT. **Tac Applicator:** GTA . **Bone Tac Case:** GB35 /GB50



HOW TO USE:

No	Product	How to Use
1	GBR Screws Motor Mount/ Driver	Connect Tac Remover Driver to a dental handpiece and tight or remove bone screws.
2	Tac Applicator	With a tip of Tac Applicator, grasp Bone Tac and place it on the surgical site.
3	Mallet	Insert Bone Tac into the surgical site by malleting Tac Applicator.


• GBR Screw driver

Type	Size	Code
	20.0 mm	MMGB20
	26.0 mm	MMGB26

- ; VVSZ

; WL		* VKL
		. ; (
		. 4 ;

-) VUL ; HJ

; WL		; WL	* VKL
		MVY ; HJ	TT .)
		MVY ; HJ	TT .)

; ; 69 (., (5+ ; 9(5:769 ; (; 065 * 65+0 ; 065 :!
 ; OL KL] PJJL ZOV\SK IL RLW[K\YPUN Z[VYHNL HUK [
 WSHJL 2LLW [OL KL] PJJL V\[VM KPYLJ[Z\USPNO[
 KL] PJJL

4 (; , 90 (:3[HPUSLZ(8\T[PLU[PSLT]HS J V ; W P S Y T P L [Y

> (9505.!

- -VY LMMPJPLU[HUK ZHML \ZL VM [OPZ KL] PJJL
HUK OH]L LUV\NO L_WLYPLUJL PU Z\YNPJHS VWL
PUZ[Y\TLU[Z
- ; OL Z\YNLVU PZ YLZWVUZPISL MVY THRPUN YLHZV
PUZ[Y\TLU[HUK Z\YNPJHS [LJOUXP\L [V \ZL MVY
JVUKP[PVU VM PUKP]PK\HS WH[PLU[

CAUTIONS:

- The surgeon should be well acquainted with suitable precautions against problems that may occur during an operation.
- Excessive drilling can cause injury during surgical operation.

PRECAUTIONS:

- The surgeon should inform the patient of the risks associated with surgery.
- Inspect each device to ensure they are not damaged.
- All instruments must be sterilized prior to use.
- The device should be used by sufficiently trained surgeon.
- The device should not be modified or used against the [Indications for Use] stated in this document.

POSSIBLE ADVERSE EFFECTS:

Allergy to metallic material, inflammation and/or infection caused by carelessness.

CONTRAINDICATION:

- Do not use for the patients with active or suspected infection, immune deficiency history, steroid treatment, bleeding disorder, endocrine diseases, osteomyelitis or other acute diseases and the patients receiving radiation therapy on the cranium or having gingival diseases or severe oral hygiene problems unless the surgeon make reasonable judgement that the surgery with the device is safe for the patients under appropriate control.
- Do not use for any patient who is not suitable for surgery including the patients with allergy to metallic material.

CLEANING & STERILIZATION:

The subject devices, surgical instruments, are reusable after cleaning and sterilizing. They must be cleaned before reuse. Following these recommended instructions for cleaning.

1) Cleaning:

- ① Disassemble the device where possible.
- ② Remove body fluids or blood with flowing water.
- ③ Cleanse the device thoroughly using suitable brushes in flowing water until foreign substances are removed. (do not use metal brushes or steel wool).
- ④ Cleanse the device with surgical instrument cleaner according to the instructions provided by the manufacturer of the cleaner.
- ⑤ Rinse the device in flowing water or distilled water for 20 seconds three times.
- ⑥ Air-dry or wipe the device with clean cloth and alcohol to prevent from water spot.

2) Pre-Cleaning:

- ① Follow the same instructions from ④ to ⑥ of Cleaning.

3) Sterilization:

Dental Surgical Instruments is supplied **NON-STERILE**.

They must be sterilized by the end user after and prior to use.

Only legally marketed, FDA-cleared sterilization barriers (e.g. wraps or pouches) should be used.

PACKAGING:

The device is packaged in KIT.

This device is manufactured made for GDT Implants.

Consult instructions for use



Manufacturer



Caution



Reference / Article number



Non-sterile product



Batch Code



Temperature limitation



Date of manufacture



Keep away from rain



Symbol for "Use by
Prescription only"

Rx Only

Keep away from sunlight



The manufacturer is not responsible for any loss of quality caused by the failure to comply with terms of transportation, storage and use established by the manufacturer for this product. Responsibility for the use of the material for purposes other than those specified by the manufacturer falls on the user.